

MxP® Quant 1000 kit requirements

This document specifies additional lab equipment and chemicals that are required but not included with the kit. Before starting with the kit, ensure that all required items are available.

LC-MS system

Agilent	SCIEX	Waters
Mass spectrometer		
6495C Triple Quadrupole LC/MS System	5500+ series (QTRAP or Triple Quad) with TurboV™ ion source	Xevo TQ-XS with ESI source
LC system and autosampler		
<ul style="list-style-type: none"> – Agilent 1290 Infinity I/II UHPLC system with column oven. – 800 bar injection valve (Agilent part no. 5067-6698) replaces standard 1300 bar valve – 1290 Infinity II inline filters, 0.3 µm (2 pcs., before and after injection valve, Agilent part no. 5067-6189) – 96-well plate autosampler with temperature control (4 °C) – Injection volume range: 2 – 20 µL 	<ul style="list-style-type: none"> – Standard flow UHPLC system with column oven. – 96-well plate autosampler with temperature control (4 °C) – Injection volume range: 2 – 20 µL 	<ul style="list-style-type: none"> – Waters ACQUITY UPLC® system with sample manager, solvent manager, and column oven: <ul style="list-style-type: none"> – Classic (loop-based) – I-Class (FTN) – H-Class (FTN) – Injection volume range: 2 – 20 µL FTN samplers require extension loop of 50 µL or needle size of 30 µL!
NanoLC and microLC are not supported!		
Tubing		
<ul style="list-style-type: none"> – Recommended between injector and analytical column: steel capillary – Recommended between analytical column and mass spectrometer: red PEEK Tubing 1/16" OD x 0.005" ID 		
Condition		
<ul style="list-style-type: none"> – Instrument is well maintained and serviced on a regular basis according to manufacturer's recommendations – MS tuned and calibrated in both positive <u>and</u> negative ion modes – LC-MS system must be free of ion pairing reagents 		

Column components (available as package or individually)

Component	Requirement
Column package	MxP® Quant 1000 kit column (1 pc.) + precolumn (3 pcs.) biocrates part number 22558
Analytical column	MxP® Quant 1000 kit column biocrates part number 22565
Precolumn	MxP® Quant 1000 kit precolumn biocrates part number 22572

Laboratory equipment

Equipment	Requirements (may be available from local distributor)	
	Nitrogen evaporator for 96-well plates	Pressure manifold for 96-well plates
Nitrogen evaporator* or pressure manifold* – Requires nitrogen supply – Must be in a fume hood – No heater required	Examples: <ul style="list-style-type: none"> – Porvair blowdown evaporator MiniVap® – Techne (FSC496D) – Sample concentrator from BenchTop Lab Systems (96 samples, BT1604) – VLM evaporators – Organomation MICROVAP microplate evaporator – Biotage TurboVap® 96 	Examples: <ul style="list-style-type: none"> – Porvair UltraPPM LITE – Waters positive pressure-96 processor (186006961) – Biotage® PRESSURE+ 96 manifold (PPM-96) – CEREX® system 96 processor (288-0001) – Agilent positive pressure manifold 96 processor (PPM-96) – TECAN Resolvex® M10 96
Centrifuge	Must be able to centrifuge 96-well plates of 5 cm height at 500 x g	Not required when a pressure manifold is used.
Shaker	Any model with adjustable speed (450 – 1200 rpm) including a tray for plates and vials. Recommended: Eppendorf MixMate® or ThermoMixer®	
Pipettes	<ul style="list-style-type: none"> – Repeater, e.g. Eppendorf Multipette® E3 (with 2.5 and 10 mL tips) or similar electronic model – Single channel: volume range 10 µL – 1000 µL – 8-Channel: 10 µL – 150 µL 	
Vortexer	Any model	
Balance	Accuracy < 1 mg	
Solvent bottles	50 – 1000 mL	

* See document “Technical guide-Nitrogen evaporators and pressure manifolds” for details.

Urine extension (optional)

Only for urine sample analysis
biocrates part number 22589

Solvents and chemicals

Solvents and chemicals	Purity
Methanol, water, acetonitrile, isopropanol	LC-MS grade
Formic acid	LC-MS grade (e.g. Honeywell Fluka™ 56302-50 mL) Fresh! New bottle or opened within past 6 months
Phenyl isothiocyanate (PITC)	99%, for protein sequencing (e.g. Sigma-Aldrich 317861) Fresh! New bottle or opened within past 6 months
Pyridine	>99% (e.g. Acros Organics 131780500, 50 mL) Fresh! New bottle or opened within past 6 months
Triethylamine	>99% (e.g. Sigma-Aldrich 90335-100 mL) Fresh! New bottle or opened within past 6 months
Ammonium acetate	LC-MS grade
Phosphate buffered saline (PBS)	p.a. grade (e.g. Sigma-Aldrich P4417)

Software

Agilent	SCIEX	Waters
MS software		
MassHunter version 10.1 or later	Analyst® version 1.7.3 or later SCIEX OS version 3.4 or later	MassLynx® version 4.2 or later
biocrates WebIDQ workflow manager		
Subscriptions available through biocrates webshop or local distributor:		
– WebIDQ basic subscription (4 GB storage space)		biocrates part number 22336.1
– WebIDQ core subscription (8 GB storage space)		biocrates part number 21629.3
– WebIDQ core+ subscription (50 GB storage space)		biocrates part number 21926.3
– WebIDQ on-premises subscription		biocrates part number 21636.3
For detailed system requirements please visit: https://shop.biocrates.com/faq/What-are-the-system-requirements-for-WebIDQ		